

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 02/24/2022

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:32
INTERNAL STANDARD	VERIFIED	20:33
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:33
BLANK TEST	0.00	20:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:34
BLANK TEST	0.00	20:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:35
BLANK TEST	0.00	20:36

