

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 03/23/2022

OPERATOR NAME:
KRISTY N ALLEN
Permit #: 49726
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:27
INTERNAL STANDARD	VERIFIED	23:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	23:28
BLANK TEST	0.00	23:29
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:29
BLANK TEST	0.00	23:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	23:30
BLANK TEST	0.00	23:31

