

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 05/02/2022

WFO

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:18
INTERNAL STANDARD	VERIFIED	18:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:18
BLANK TEST	0.00	18:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:19
BLANK TEST	0.00	18:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:20
BLANK TEST	0.00	18:21

