

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 07/27/2022

OPERATOR NAME:
JAY V SORRENTINO
Permit #: 38962
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:29
INTERNAL STANDARD	VERIFIED	03:29
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:29
BLANK TEST	0.00	03:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:31
BLANK TEST	0.00	03:31
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:32
BLANK TEST	0.00	03:33

