

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

WFO

Date: 08/10/2022

OPERATOR NAME:  
URIAH J PORTNER  
Permit #: 50824  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:07
INTERNAL STANDARD	VERIFIED	19:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:07
BLANK TEST	0.00	19:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:08
BLANK TEST	0.00	19:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:10
BLANK TEST	0.00	19:11

