

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/07/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22190

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 16:14 |
| INTERNAL STANDARD | VERIFIED | 16:14 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.10 | 16:14 |
| BLANK TEST | 0.00 | 16:15 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.10 | 16:16 |
| BLANK TEST | 0.00 | 16:16 |
| REFERENCE STANDARD TEMP | 34.03c | |
| REFERENCE STANDARD | 0.10 | 16:17 |
| BLANK TEST | 0.00 | 16:18 |

