

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 10/26/2022

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:25
INTERNAL STANDARD	VERIFIED	00:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:25
BLANK TEST	0.00	00:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:26
BLANK TEST	0.00	00:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:28
BLANK TEST	0.00	00:29

