

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 11/18/2022

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22260

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:46
INTERNAL STANDARD	VERIFIED	18:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:46
BLANK TEST	0.00	18:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:48
BLANK TEST	0.00	18:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:49
BLANK TEST	0.00	18:50

