

TECHNICIAN TEST

WFO

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 03/01/2022

Nominal = 0.100 g/210L

BLANK TEST	0.000	10:12
INTERNAL STANDARD	VERIFIED	10:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.104	10:12
BLANK TEST	0.000	10:13
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.104	10:13
BLANK TEST	0.000	10:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.104	10:14
BLANK TEST	0.000	10:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.104	10:16
BLANK TEST	0.000	10:16

Average = 0.1040  
Std Dev = 0.0000

