

SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 05/03/2023

WFO

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:29
INTERNAL STANDARD	VERIFIED	02:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:30
BLANK TEST	0.00	02:31
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	02:31
BLANK TEST	0.00	02:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:33
BLANK TEST	0.00	02:33

