

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 06/01/2023

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:03
INTERNAL STANDARD	VERIFIED	00:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:04
BLANK TEST	0.00	00:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:05
BLANK TEST	0.00	00:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:06
BLANK TEST	0.00	00:07

