

SOLUTION CHANGE

-----  
ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197  
-----

Date: 11/01/2023

OPERATOR NAME:  
STEVEN P EILERS  
Permit #: 49321  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:18
INTERNAL STANDARD	VERIFIED	01:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:18
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:19
BLANK TEST	0.00	01:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:21
BLANK TEST	0.00	01:21

