

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 02/01/2023

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:13
INTERNAL STANDARD	VERIFIED	02:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:13
BLANK TEST	0.00	02:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:15
BLANK TEST	0.00	02:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:16
BLANK TEST	0.00	02:17

WFO

