

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 11/15/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:57
INTERNAL STANDARD	VERIFIED	01:57
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:58
BLANK TEST	0.00	01:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:59
BLANK TEST	0.00	02:00
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:00
BLANK TEST	0.00	02:01

