

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 06/14/2023

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:49
INTERNAL STANDARD	VERIFIED	01:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:50
BLANK TEST	0.00	01:51
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:51
BLANK TEST	0.00	01:52
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:52
BLANK TEST	0.00	01:53

