

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 07/05/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:00
INTERNAL STANDARD	VERIFIED	01:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:01
BLANK TEST	0.00	01:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:02
BLANK TEST	0.00	01:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:03
BLANK TEST	0.00	01:04

