

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

WFO

Date: 07/12/2023

OPERATOR NAME:  
JOHN P CUSACK  
Permit #: 35563  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:55
INTERNAL STANDARD	VERIFIED	19:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:55
BLANK TEST	0.00	19:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:57
BLANK TEST	0.00	19:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	19:58
BLANK TEST	0.00	19:59

