

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 07/26/2023

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:36
INTERNAL STANDARD	VERIFIED	22:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:36
BLANK TEST	0.00	22:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:38
BLANK TEST	0.00	22:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	22:39
BLANK TEST	0.00	22:40

