

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFB

Date: 09/27/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:04
INTERNAL STANDARD	VERIFIED	18:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:05
BLANK TEST	0.00	18:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:06
BLANK TEST	0.00	18:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:07
BLANK TEST	0.00	18:08

