

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 01/18/2023

OPERATOR NAME:
NOAH A DANQUER
Permit #: 50812
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 22390

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 00:19 |
| INTERNAL STANDARD | VERIFIED | 00:19 |
| REFERENCE STANDARD TEMP | 34.03c | |
| REFERENCE STANDARD | 0.10 | 00:20 |
| BLANK TEST | 0.00 | 00:20 |
| REFERENCE STANDARD TEMP | 34.03c | |
| REFERENCE STANDARD | 0.10 | 00:21 |
| BLANK TEST | 0.00 | 00:22 |
| REFERENCE STANDARD TEMP | 34.03c | |
| REFERENCE STANDARD | 0.10 | 00:22 |
| BLANK TEST | 0.00 | 00:23 |

