

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 11/29/2023

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:42
INTERNAL STANDARD	VERIFIED	02:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:42
BLANK TEST	0.00	02:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:43
BLANK TEST	0.00	02:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:44
BLANK TEST	0.00	02:45

