

Infrared Instrument
Weekly Calibration Verification Test Record

WRO

Department: 060

Instrument Serial #: 100197

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
6/1/23	23090	.10	.10	.10	MP4192	34.02	34.02	34.02	Cusack	Y
6/8/23	23090	.10	.10	.10	MP4192	34.03	34.02	34.03	Eilers	N
6/14/23	23090	.10	.10	.10	MP4192	34.03	34.03	34.03	Eilers	N
6/21/23	23090	.10	.10	.10	MP4192	34.02	34.02	34.02	Osika	N
7/1/23	23090	.10	.10	.10	MP4192	34.02	34.02	34.02	Eilers	Y
7/3/23	23090	.10	.10	.10	MP4192	34.02	34.02	34.02	Eilers	N
7/12	23090	.10	.10	.10	MP4192	34.02	34.02	34.02	Cusack	N
7/19/23	23090	.10	.10	.10	MP4192	34.02	34.02	34.02	Eilers	N
7/26/23	23090	.10	.10	.10	MP4192	34.02	34.02	34.02	Osika	N
8/1/23	23190	.10	.10	.10	MP4192	34.02	34.02	34.02	Osika	Y
8/10/23	23190	.10	.10	.10	MP4192	34.02	34.02	34.02	Osika	N
8/16/23	23190	.10	.10	.10	MP4192	34.02	34.02	34.02	Osika	N

* - Please enter a √ to denote simulator solution change taking place