

WFO

**Infrared Instrument
Weekly Calibration Verification Test Record**

Department: OC50

Instrument Serial #: 100197

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
3/8/23	22420	34.03	34.03	34.03	53215	.10	.10	.10	Cusack	N
3/29/23	22420	34.03	34.03	34.03	53215	.10	.10	.10	Osika	Y
4/1/23	23030	34.03	34.02	34.03	53215	.10	.10	.10	Osika	Y
4/6/23	23030	34.03	34.03	34.03	53215	.10	.10	.10	Osika	N
4/12/23	23030	34.02	34.02	34.02	53215	.10	.10	.10	Osika	N
4/19/23	23030	34.03	34.03	34.03	53215	.10	.10	.10	Eilers	N
4/27/23	23030	34.03	34.03	34.02	53215	.10	.10	.10	Eilers	N
5/3/23	23030	34.02	34.02	34.03	53215	.10	.10	.10	Eilers	Y
5/10/23	23030	34.02	34.02	34.02	53215	.10	.10	.10	Jorrevino	N
5/10/23	23030	34.02	34.02	34.02	53215	.10	.10	.10	Eilers	N
5/17/23	23030	34.02	34.02	34.02	53215	.10	.10	.10	Eilers	N
5/24/23	23030	34.03	34.03	34.03	53215	.10	.10	.10	Eilers	N

* - Please enter a Y to denote simulator solution change taking place