

SOLUTION CHANGE

WFD

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100197

Date: 01/01/2024

OPERATOR NAME:
 JASON L OSIKA
 Permit #: 35567
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:56
INTERNAL STANDARD	VERIFIED	21:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:57
BLANK TEST	0.00	21:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:58
BLANK TEST	0.00	21:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:59
BLANK TEST	0.00	22:00

