

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 03/07/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:26
INTERNAL STANDARD	VERIFIED	01:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:26
BLANK TEST	0.00	01:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:27
BLANK TEST	0.00	01:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:28
BLANK TEST	0.00	01:29

