

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 04/04/2024

OPERATOR NAME:
JOHN P CUSACK
Permit #: 35563
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24040

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:07
INTERNAL STANDARD	VERIFIED	02:07
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:08
BLANK TEST	0.00	02:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:09
BLANK TEST	0.00	02:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:10
BLANK TEST	0.00	02:11

