

SUPERVISOR TEST

WFO

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100197

Date: 01/17/2024

OPERATOR NAME:
 JASON L OSIKA
 Permit #: 35567
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:14
INTERNAL STANDARD	VERIFIED	01:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:15
BLANK TEST	0.00	01:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:16
BLANK TEST	0.00	01:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:17
BLANK TEST	0.00	01:18

